

Date	Time	Accmulated Exposure Clock Set Power On
3/22/2016	9:58:12	0
3/22/2016	10:08:12	0
3/22/2016	10:18:12	0
3/22/2016	10:28:12	0
3/22/2016	10:38:12	0
3/22/2016	10:48:12	0
3/22/2016	10:58:12	0
3/22/2016	11:08:12	0
3/22/2016	11:18:12	0
3/22/2016	11:28:12	0
3/22/2016	11:38:12	0
3/22/2016	11:48:12	0
3/22/2016	11:58:12	0
3/22/2016	12:08:12	0
3/22/2016	12:18:12	0
3/22/2016	12:28:12	0
3/22/2016	12:38:12	0
3/22/2016	12:48:12	0
3/22/2016	12:58:12	0
3/22/2016	13:08:12	0
3/22/2016	13:18:12	0
3/22/2016	13:28:12	0
3/22/2016	13:38:12	0
3/22/2016	13:48:12	0
3/22/2016	13:58:12	0
3/22/2016	14:08:12	0
3/22/2016	14:18:12	0
3/22/2016	14:28:12	0
3/22/2016	14:38:12	0
3/22/2016	14:48:12	0
3/22/2016	14:58:12	0
3/22/2016	15:08:12	0
3/22/2016	15:18:12	0
3/22/2016	15:28:12	0
3/22/2016	15:38:12	0
3/22/2016	15:48:12	0
3/22/2016	15:58:12	0
3/22/2016	16:08:12	0
3/22/2016	16:18:12	0
3/22/2016	16:28:12	0
3/22/2016	16:38:12	0
3/22/2016	16:48:12	0
3/22/2016	16:58:12	0
3/22/2016	17:08:12	0

3/22/2016	17:18:12	0
3/22/2016	17:28:12	0
3/22/2016	17:38:12	0
3/22/2016	17:48:12	0
3/22/2016	17:58:12	0
3/22/2016	18:08:12	0
3/22/2016	18:18:12	0
3/22/2016	18:28:12	0.1
3/22/2016	18:38:12	0.1
3/22/2016	18:48:12	0.1
3/22/2016	18:58:12	0.1
3/22/2016	19:08:12	0.1
3/22/2016	19:18:12	0.1
3/22/2016	19:28:12	0.1
3/22/2016	19:38:12	0.1
3/22/2016	19:48:12	0.1
3/22/2016	19:58:12	0.1
3/22/2016	20:08:12	0.1
3/22/2016	20:18:12	0.1
3/22/2016	20:28:12	0.1
3/22/2016	20:38:12	0.1
3/22/2016	20:48:12	0.1
3/22/2016	20:58:12	0.1
3/22/2016	21:08:12	0.1
3/22/2016	21:18:12	0.1
3/22/2016	21:28:12	0.1
3/22/2016	21:38:12	0.1
3/22/2016	21:48:12	0.1
3/22/2016	21:58:12	0.1
3/22/2016	22:08:12	0.1
3/22/2016	22:18:12	0.1
3/22/2016	22:28:12	0.1
3/22/2016	22:38:12	0.1
3/22/2016	22:48:12	0.1
3/22/2016	22:58:12	0.1
3/22/2016	23:08:12	0.1
3/22/2016	23:18:12	0.1
3/22/2016	23:28:12	0.1
3/22/2016	23:38:12	0.1
3/22/2016	23:48:12	0.1
3/22/2016	23:58:12	0.1
3/23/2016	0:08:12	0.1
3/23/2016	0:18:12	0.1
3/23/2016	0:28:12	0.1
3/23/2016	0:38:12	0.1
3/23/2016	0:48:12	0.1
3/23/2016	0:58:12	0.1

3/23/2016	1:08:12	0.1
3/23/2016	1:18:12	0.1
3/23/2016	1:28:12	0.1
3/23/2016	1:38:12	0.1
3/23/2016	1:48:12	0.1
3/23/2016	1:58:12	0.2
3/23/2016	2:08:12	0.2
3/23/2016	2:18:12	0.2
3/23/2016	2:28:12	0.2
3/23/2016	2:38:12	0.2
3/23/2016	2:48:12	0.2
3/23/2016	2:58:12	0.2
3/23/2016	3:08:12	0.2
3/23/2016	3:18:12	0.2
3/23/2016	3:28:12	0.2
3/23/2016	3:38:12	0.2
3/23/2016	3:48:12	0.2
3/23/2016	3:58:12	0.2
3/23/2016	4:08:12	0.2
3/23/2016	4:18:12	0.2
3/23/2016	4:28:12	0.2
3/23/2016	4:38:12	0.2
3/23/2016	4:48:12	0.2
3/23/2016	4:58:12	0.2
3/23/2016	5:08:12	0.2
3/23/2016	5:18:12	0.2
3/23/2016	5:28:12	0.2
3/23/2016	5:38:12	0.2
3/23/2016	5:48:12	0.2
3/23/2016	5:58:12	0.2
3/23/2016	6:08:12	0.2
3/23/2016	6:18:12	0.2
3/23/2016	6:28:12	0.2
3/23/2016	6:38:12	0.2
3/23/2016	6:48:12	0.2
3/23/2016	6:58:12	0.2
3/23/2016	7:08:12	0.2
3/23/2016	7:18:12	0.2
	Power On	
	Power On	
	Power On	
	Power On	
	Power On	
6/7/2016	11:05:14	0
6/7/2016	11:15:14	0
	Power On	
6/7/2016	12:00:51	0

6/7/2016	12:10:51	0
6/7/2016	12:20:51	0
6/7/2016	12:30:51	0
6/7/2016	12:40:51	0
6/7/2016	12:50:51	0
6/7/2016	13:00:51	0
6/7/2016	13:10:51	0
6/7/2016	13:20:51	0
6/7/2016	13:30:51	0
6/7/2016	13:40:51	0
6/7/2016	13:50:51	0
6/7/2016	14:00:51	0
6/7/2016	14:10:51	0
6/7/2016	14:20:51	0
6/7/2016	14:30:51	0

Dead Time		5.00E-05
Calibration Factor		1.35
Firmware Ver		1.17
Units	mR	
Dose Warn	0 mR	
Dose Alarm	50 mR	
Rate Warn	0 mR/h	
Rate Alarm	500 mR/h	
Sensor Fail	10 min	
Logging Freq	10 min	
Calibration Factor		1.35
Dead Time		5.00E-05