

Always keep the main gaskets clean and free from sand and dirt. Contamination can lead to the case leaking and damage to the sealing surfaces. The ruggedized case is not designed for underwater use.



Open all 4 (four) side-closure latches



Release the straps



Install Inspector, centering over screen opening. Reattach the straps to hold the instrument in place



Replace foam and press case halves together



Hook and close all latches. You are now ready to use your ruggedized Inspector



Lift hard shell window to access the Inspector controls.

